

C36 VOCs IN SOIL

1.0 Sample Reception

- 1.1 Provided as 4 x 8 g soil and 20 mL methanol in 40mL VOA vials.
- 1.2 All breakages and shortages must be reported within 24 hours of sample receipt.
- 1.3 Store samples at 4°C ± 3°C.
- 1.4 Check that all the parameters for which you are registered are correctly identified on the web data entry report page.
- 1.5 Inquiries regarding samples and their shipment may be directed to:

PT Non-conformances

Phenova
Tel: (866) 942-2978
Fax: (866) 283-0269
Email: TyG@phenova.com

cc: PT Canada, Program Administrator
email: programadmin@PTcanada.org
cc: Ken Middlebrook, PT Canada
email: kmiddlebrook@PTcanada.org

Inquiries must be made by email only. Use the enclosed Nonconformance Form (see reverse) when notifying PT Canada of a problem with the samples. Please include your PT Canada laboratory number on all correspondence.

2.0 Sample Analysis

- 2.1 Perform the analysis on these samples assuming that each sample contains 8 g dry weight of soil and 20 mL of methanol.
- 2.2 VOC compounds are in the concentration range typical for contaminated soil.
- 2.3 Proceed with testing using the routine analytical method identified in your PT Canada application.

3.0 Reporting Results

- 3.1 Results must be reported by midnight of the study deadline (see the General Proficiency Testing Information sheet for details).
- 3.2 Report results on a dry weight basis.

4.0 Safety

- 4.1 The PT samples are designed for use by laboratory professionals familiar with environmental samples and potentially hazardous materials.

PT SAMPLE NON-CONFORMANCE FORM

Attn: PT non-conformances

Study Number:

ENSURE THAT SAMPLES RECEIVED MATCH REPORT FORMS

1 - Laboratory Information

Contact Name:

Laboratory Name

Laboratory Address

Contact Telephone #

Contact Facsimile #

Contact e-mail:

2 - Sample Details

Date & Time of Arrival(YYYY,MM,DD,HH:MM):

Tracking Number:

Test Groups Received (e.g. C1, C2 etc.):

Number of Boxes:

3 - Description of Nonconformance

4 - Requested Action

5 - PT Provider Notes